

- 2- Triangular bandages
- 2- Sterile burn sheets
- 3-Vaseline Gauze
- 10- Alcohol preps
- 1- Bandage shears
- 1- Pen light
- 1- Multilevel stretcher with all patient securing straps as recommended by the stretcher's manufacturer and at least 2 sets of clean sheets and blankets
- 1- Mast pants (Optional)
- 1ea- Air-Qsp3G Airway Devices, sizes 0, 0.5, 1.0, 2.0, 3, 4, and 5
- 1- Puncture resistant sharps container
- 1- Emergency Response Guidebook, most current edition
- 1set- Emergency warning devices (at least 3 emergency triangles)
- 10- Protective, non-porous gloves
- 2- Medical eye protection
- 5- N-95 or greater protective Masks
- 2- Medical protective gowns or equivalent
- 2- Personal cleansing supplies
- 5- Biohazard bags
- 1- Working flashlight
- 1- Fire extinguisher
- 1- No smoking signs in the cab and patient compartment
- 1- Thermometer (may be oral, tympanic, or skin monitoring)
- 1- Mechanical CPR Compression Device (Optional). If carried, services are to follow manufacturer's recommendation assuring compliance with current AHA CPR guidelines
- 4- Cold packs or equivalent

ALS & ALS-CAPABLE UNITS

ALL BLS EQUIPMENT AND

- Advanced airway equipment including equipment required to perform endotracheal intubation including but not limited to:
 - ◆ 1-Laryngoscope handle with appropriate spare batteries
 - ◆ 1ea- Laryngoscope blades (Miller 1, 2, 3, 4 and Mac 1, 2, 3, 4)
 - ◆ 1ea- Extra laryngoscope bulbs (small and large) (extra bulbs not required for fiber optic laryngoscope sets)
 - ◆ 2ea- Endotracheal tubes (adult, pediatric sizes, 4 thru 9.0)
 - ◆ 1- ET tube holder
 - ◆ 1ea- Stylette (adult, pediatric sizes)
- 1ea- ET tube introducer (i.e., Bougie) (adult and pediatric)
- 2- Soft suction catheters for suctioning of ET tube
- 1- Carbon Dioxide monitor or detector (if a Carbon Dioxide detector (i.e., Easy Cap) is solely utilized an adult and pediatric must be stocked)
- 1ea- Magill forceps (adult and pediatric)
- Equipment & supplies to establish intravenous infusions for adult & pediatric patients to include but not limited to:
 - ◆ 5ea- IV catheters (16ga, 18ga, 20ga, 22ga)

- ◆ 4-1000cc Normal Saline or equivalent volume
- ◆ 4-1000cc Lactated Ringers or equivalent volume
- ◆ 4-Macro IV tubing
- ◆ 2-Micro IV tubing
(If dial-a-flow tubing is used then a total of 6 IV tubes are required)
- ◆ 4-Securing devices (nonporous tape or Veni-guard devices are acceptable)
- 1- EZ IO Driver with spare batteries (EZ IO Drivers with the non-replaceable battery need not to have a spare)
- 1- EZ IO PD needle
- 1- Adult EZ-IO needle
- 1- EZ IO LD needle (Optional)
- 1- EZ-STABILIZER (Optional)
- 2ea- Blue, Red, Purple, and Green blood specimen tubes or their equivalent
- 2- 14ga X 2" or longer IV catheters for pleural decompression
- 1- Morgan Lenses (Optional)

MICU & MICU-CAPABLE UNITS

ALL BLS, ALS EQUIPMENT AND

- 1- Copy of current, signed SPEMS treatment protocols for EMT-Paramedic
- 1- Cardiac monitor/defibrillator with 2 sets of defibrillator pads/paddles to accommodate adult patients and 2 sets of defibrillator pads/paddles to accommodate pediatric patients (if not already stocked at the ALS or BLS level). A charged spare battery must accompany the unit as well as the one powering the unit. However, an alternative power source may take the place of the spare battery.
- 10- EKG electrodes
- 1- Commercial cryothyrotomy kit **OR** a preassembled cryothyrotomy kit including but not limited to:
 - ◆ 1-Scalpel
 - ◆ 1-Hemostat
 - ◆ 1-Gauze pad
 - ◆ 1-Betadine swab
- CPAP supplies/equipment (optional) (CPAP equipment must be approved by Medical Director)
- 1- 100cc or 250cc Normal Saline (For IV mix)
- 1- 100cc D5W (optional)

BLS NEONATAL EQUIPMENT

- 1- Sealed OB kit with non-porous infant insulator, umbilical cord clamps and bulb aspirator
- 1-Broselow Pediatric Emergency Tape or equivalent
- 1-Infant Bag valve mask (BVM)
- 1-#1 Oral airway

ALS NEONATAL EQUIPMENT

In addition to the BLS Neonatal Equipment list, all ALS or ALS Capable units must be equipped with:

- 1- Miller size 0 laryngoscope blade with extra bulb (extra bulb not required for fiber optic laryngoscope blade)
- 2ea- Endotracheal tubes (sizes, 2.5, 3.0, and 3.5)
- 5- 24ga IV catheters